FLOW CHART

Below is a flowchart showing how to navigate through all menus by pressing front buttons.



DISPLAY ABBREVIATIONS

ALR1	Alarm 1 Status		
OFF	Alarm 1 set Off	**ON**	Alarm 1 set On
A1Md	Alarm 1 Mode		
A1LO	Alarm 1 Low	**A1HI**	Alarm 1 High
A1LH	Alarm 1 Low/High		
LO-1	Alarm 1 Low	**-999.. ..9999**	Alarm 1 Low Value
HI-1	Alarm 1 High	**-999.. ..9999**	Alarm 1 High Value
A1CR	Display color when Alarm 1 triggered		
GRN	Green Color	**REd**	Red Color
AMbR	Amber Color		
ALR2	Alarm 2 Status		
OFF	Alarm 2 set Off	**ON**	Alarm 2 set On
A2Md	Alarm 2 Mode		
A2LO	Alarm 2 Low	**A2HI**	Alarm 2 High
A2LH	Alarm 2 Low/High		
LO-2	Alarm 2 Low	**-999.. ..9999**	Alarm 2 Low Value
HI-2	Alarm 2 High	**-999.. ..9999**	Alarm 2 High Value
A2CR	Display color when Alarm 2 triggered		
GRN	Green Color	**REd**	Red Color
AMbR	Amber Color		
OUt	Alarm Latched/Unlatched selection		
LAtC	Latched	**UNLA**	Unlatched
NO.CR	Display Color in Normal condition		
GRN	Green Color	**REd**	Red Color
AMbR	Amber Color		
MOdE	Data Flow Mode		
HOSt	Host Mode	**SLAV**	Slave Mode
bAUd	Baud Rate	**300.. ..19200**	Baud Rate Value
FORM	Data Format		
7O1	7 Bit, Odd, 1 Stop Bit	**7E1**	7 Bit, Even, 1 Stop Bit
8N1	8 Bit, No parity, 1 Stop Bit		
COMM	Communication Standard		
232	RS-232 Standard	**485**	RS-485 Standard
AddR	Device Address	**0000.. ..0099**	Address Value
INtF	Interface Device		
dRNt	DRN with Temperature Input	**dRNP**	DRN with Process Input
Miscellaneous:			
PEAk	Peak Value	**VALL**	Valley Value
PROC	Process Value	**RUN**	Run Mode
OVLd	Input Overload	**StOR**	Stored Message

Note

1. In **Slave** Mode the Big Display will wait for commands and data from the Serial Bus.
2. In **Host** Mode the Big Display will send data automatically and continuously into the Serial Bus.
3. When used in **RS-485** Mode, the device must be accessed with an appropriate **Address** Value.
4. **Latched** Mode: Alarm remains latched until reset. To reset already latched alarm select any menu items and then press “up” or “down” button.

SPECIFICATION

Temperature Stability:
50 ppm/°C

Display:
4-digit, 7-segment LED, 57.2mm (2.25”) with red, green and amber programmable colors.

Alarm:
Alarm 1 & 2 programmable, Latch/Unlatch, High, Low, High/Low

SERIAL INTERFACE

Communication Standard:
RS-485, RS-422 or RS-232

Transfer speed (Baud rate):
300, 600, 1200, 2400, 4800, 9600, 19200 bps

Data Format:
7O1-7 bit, Odd, 1 stop bit, 7E1- 7 bit, even, 1 stop bit
8N1 – 8 bit, No parity, 1 stop bit

Multi-Point Address (RS-485):
0 to 199

Flow Control:
No Flow control

Screw terminals for RS-232/485/422 interface

Power Supply:
100-240 Vac ±10%, 50/60 Hz, 22.5 W

Operating Temperature:
0 to 40°C

Storage Temperature:
-20 to 60°C

Relative Humidity:
0 to 85%

Protection:
NEMA-4x (IP65)

Dimensions:
298 L x 137 W x 73 D mm
(11.75” x 5.375” x 2.875”)

Panel Cutout:
279.4 L x 116.8 W mm
(11.00” L x 4.60” W)

Weight:
1,360 g (3 lbs)

Approvals:
CE per EN61010-1:2001

WARNING: These products are not designed for use in, and should not be used for, patient-connected applications.

This device is marked with the international caution symbol. It is important to read the Setup Guide before installing or commissioning this device, as the guide contains important information relating to safety and EMC.

It is the policy of OMEGA to comply with all worldwide safety and EMC/EMI regulations that apply. OEMGA is constantly pursuing certification of its products to the European New Approach Directives. OMEGA will add the CE mark to every appropriate device upon certification.

The information contained in this document is believed to be correct, but OMEGA Engineering, Inc. accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.

TRADEMARK NOTICE:

omega.com®, OMEGA®, and iTM are Trademarks of OMEGA ENGINEERING, INC.

M4191/0405

OPERATION MANUAL

iSeries®

iLD24-C2
Big Remote Display with RS-232 Input

omega.com®
OMEGA®

OMEGAnet® On-Line Service www.omega.com	**Internet e-mail info@omega.com**

Servicing North America:

USA: ISO 9001 Certified
One Omega Drive, P.O. Box 4047
Stamford CT 06907-0047
TEL: (203) 359-1660 FAX: (203) 359-7700
e-mail: info@omega.com

Canada:
976 Bergar
Laval (Quebec) H7L 5A1
TEL: (514) 856-6928 FAX: (514) 856-6886
e-mail: info@omega.ca

For immediate technical or application assistance:

USA and Canada:
Sales Service: 1-800-826-6342 / 1-800-TC-OMEGA®
Customer Service: 1-800-622-2378 / 1-800-622-BEST®
Engineering Service: 1-800-872-9436 / 1-800-USA-WHEN®

Mexico and Latin American:
TEL: (001)800-TC-OMEGA® FAX: (001) 203-359-7807
En Español: (001) 203-359-7803
e-mail: espanol@omega.com

Servicing Europe:

Benelux:
Postbus 8034, 1180 LA Amstelveen, The Netherlands
TEL: +31 20 3472121 FAX: +31 20 6434643
Toll Free in Benelux: 0800 0993344
e-mail: sales@omegaeng.nl

Czech Republic:
Frystatska 184, 733 01 Karviná
TEL: +420 59 6311899 FAX: +420 59 6311114
e-mail: info@omegashop.cz

France:
11, rue Jacques Cartier, 78280 Guyancourt
TEL: +33 1 61 37 29 00 FAX: +33 1 30 57 54 27
Toll Free in France: 0800 466 342
e-mail: sales@omega.fr

Germany/Austria:
Daimlerstrasse 26, D-75392 Deckenpfronn, Germany
TEL: +49 7056 9398-0 FAX: +49 7056 9398-29
Toll Free in Germany: 0800 639 7678
e-mail: info@omega.de

United Kingdom: ISO 9002 Certified
One Omega Drive
River Bend Technology Centre
Northbank, Irlam Manchester M44 5BD United Kingdom
TEL: +44 161 777 6611 FAX: +44 161 777 6622
Toll Free in England: 0800 488 488
e-mail: sales@omega.co.uk

DESCRIPTION:

The iLD24 is a 4-digit master/slave display providing remote readout from instruments such as programmable controllers, digital panel meters and other instruments with serial output. Communication interfaces supported are RS-232 or RS-485 standards. Both RS-232 or RS-485 are programmable through front panel buttons.

The Big Display features a large three color programmable display with the capabitity to change color every time an Alarm is triggered.

SAFETY:

- The instrument is a panel mount device protected in accordance with EN61010-1:2001.

EMC:

- Whenever EMC is an issue, always use shielded cables.
- Never run signal and power wires in the same conduit.
- Use signal wire connections with twisted-pair cables.
- Install Ferrite Bead(s) on signal wire close to the instrument if EMC problems persist.

MOUNTING



Mounting Big Display Through Panel:

1. Using the panel cutout diagram shown above, cut an opening in the panel.
2. Remove six screws at the back of Big Display to remove back cover.
3. Insert the unit into the opening from the front of the panel, so the gasket seals between the bezel and the front of the panel.
4. Align back cover to Big Display and reinstall screws.

Mounting Big Display on Bail:

1. Use the Big Display template to mark the location of mounting screws on the flat surface.
2. Be sure to leave enough room around the bail (as noted on the template drawing) to allow for removal and rotation of the display.
3. The display can be rotated for the best viewing angle.

Disassembly Instruction:

Warning: Disconnect all ac power from the unit before proceeding.

1. Remove all wiring connections from the rear of the instrument, by unscrewing the power and input connectors.
2. Remove six screws at the back of the display and back cover.
3. Remove the Big Display from the panel.
4. To remove the Big Display from the bail, unscrew the two knobs at each end of the mounting brackets.

WIRING

1. Wiring RS-232 Interface.

The RS-232 standard (point-to-point) allows a single device to be connected to the Big Display using a three-wire connection (full duplex).



Device with RS-232 Pin Function	Large Remote Display	
	RJ-12	Screw Terminal
Receive (Rx)	4 (Tx)	3 (Tx)
Transmit (Tx)	3 (Rx)	2 (Rx)
Common Ground (COM)	5	1

2. Wiring RS-485 Interface.

The RS-485 standard (multipoint) allows a computer, one or more devices and Big Displays (up to 32) to be connected using a two-wire connection (half-duplex) plus a common wire to connect to the shield of the cable. It is recommended to use shielded cable with one twisted pair for EMI noise protection.



Note Connections to the computer are optional.

Computer Card or Converter Box Pin Function	Device with RS-485 Pin Function	Remote Display	
		RJ-12	Screw Terminal
A, -Tx/-Rx	-Tx/-Rx	4	3
B, +Tx/+Rx	+Tx/+Rx	3	2
COM	COM		1

3. Power Connection.

Connect the main power connections as shown in the figure below.



OPERATIONS

1. Peak Value (Display in Host Mode)

Press ▲ to request "Peak" value:
a) RS-232 Mode, will send:
*X02 (Interface DRNT), or *X03 (Interface DRNP)
b) RS-485 Mode, will send:
*01X02 (Interface DRNT), or *01X03 (Interface DRNP)

Note In the examples for RS-485 it is assumed that the device address is 01.

2. Valley Value (Display on Host Mode)

Press ▼ to request "Valley" value.
a) RS-232 Mode, will send:
*X03 (Interface DRNT), or *X04 (Interface DRNP)
b) RS-485 Mode, will send:
*01X03 (Interface DRNT), or *01X04 (Interface DRNP)

3. Process Value (Display on Host Mode)

Press ↵ to request "Process" Value.
a) RS-232 Mode, will send: *X01
b) RS-485 Mode, will send: *01X01

4. Write alphanumeric characters to the Big Display from the computer (Display in Slave Mode)

a) Single Big Display: (RS232) write 4 characters, then CR (carriage return)
b) Multiple Big Display: (RS485) write *, device address (2 digit), CR, 4 characters, then CR

5. Display Color Setup (Alarm Setup)

This menu allows the user to select the color of the display in normal conditions and when alarm is triggered. If user wants the Display to change color every time when both Alarm 1 and Alarm 2 are triggered, the Alarm values should be set in such a way that Alarm 1 is always on the top of Alarm 2 value, otherwise value of the Alarm 1 will overwrite value of Alarm 2 and Display color would not change when Alarm 2 is triggered.

Example 1:
Alarm 1 setup: "ON", Alarm Mode High "A1HI", Alarm High Value "HI-1"=400, Alarm Color "A1CR"=Amber
Alarm 2 setup: "ON", Alarm Mode High "A2HI", Alarm High Value "HI-2"=200, Alarm Color "A2CR"=Red
Normal Color: "NO.CR"=Green

Display colors change sequences:

GREEN | RED | AMBER
0 HI-2 = 200 HI-1 = 400

Example 2:
Alarm 1 setup: "ON", Alarm Mode Low "A1LO", Alarm Low Value "LO-1"=100, Alarm Color "A1CR"=Amber
Alarm 2 setup: "ON", Alarm Mode LO "A2LO", Alarm High Value "LO-2"=300, Alarm Color "A2CR"=Red
Normal Color: "NO.CR"=Green

Display colors change sequences:

AMBER | RED | GREEN
0 LO-1 = 100 LO-2 = 300

Example 3:
Alarm 1 setup: "ON", Alarm Mode Low/High "A1LH", Alarm Low Value "LO-1"=100, Alarm High Value "HI-1"=250, Alarm Color "A1CR"=Amber
Alarm 2 setup: "ON", Alarm Mode Low/High "A2LH", Alarm Low Value "LO-2"=150, Alarm High value "HI-2"=200, Alarm Color "A2CR"=Red
Normal Color: "NO.CR"=Green

Display colors change sequences:

AMBER | RED | GREEN | RED | AMBER
0 LO-1 = 100 LO-2 = 150 HI-2 = 200 HI-2 = 250

CONFIGURATION

Button Functions in Configuration Mode

↵ (MENU)	• To enter the Menu, the user must first press ↵ button. • Use this button to advance/navigate to the next menu item. The user can navigate through all the top level menus by pressing ↵. • While a parameter is being modified, press ↵ to escape without saving the parameter.
▲ (UP)	• Press the up ▲ button to scroll through submenu selections. When a numerical value is displayed press this key to increase value of a parameter that is currently being modified. • In the Run Mode pressing ▲ causes the display to flash the PEAK value several times before returning to the Run Mode. • In the top menu press ▲ causes the display to return to the Run Mode.
▼ (DOWN)	• Press the down ▼ button to scroll through submenu selections. When a numerical value is displayed press this key to decrease value of a parameter that is currently being modified. • In the Run Mode press ▼ causes the display to flash the Valley value several times before returning to the Run Mode. • In the top menu press ▼ causes the display to return to the Run Mode.
↵ (ENTER)	• Press this button to access the submenus from a Top Level Menu item. • Press this button to store a submenu selection or after entering a value – the display will flash a `StOR` message to confirm your selection.

Note x, w, z, and some punctuations are non-printable characters.